

PLEASE REFER TO IMPORTANT MESSAGES ON FINAL PAGE BEFORE APPLICATION OF ANY CLEANER/DISINFECTANT.

Sorted by Product Name:

MANUFACTURER	PRODUCT	CONCENTRATION	DATE TESTED
Ecolab®	A-456-II	Minimum 64 to 1 dilution with water	02-Dec-14
Quantum Pacific Limited	Actimax	According to Manufacturer Spec	11-Apr-06
ABC Compounding Co., Inc.	Aero Stain Away	According to Manufacturer Spec	11-Aug-16
SCIENCECompany	Ammonium Chloride	According to Manufacturer Spec	17-May-19
Milton®	Anti-Bacterial Tablets	1 Tablet:2L Water	19-Jul-12
Ecolab®	Asepti Active Liquid	According to Manufacturer Spec	20-Nov-12
Ecolab®	Asepti Active Wipes	Ready to Use	20-Nov-12
Synergy Health	Azomax™ Multi Surface Wipes	Ready to Use	15-Mar-13
Synergy Health	Azowipe™ Hard Surface Wipes	Ready to Use	15-Mar-13
PreVasive®	BAC Botanical Antimicrobial Cleaner	According to Manufacturer Spec	22-May-19
MEDCHEM®	BAK 1:750	According to Manufacturer Spec	12-May-19
Clorox®	Bleach Wipes (Sodium Hypochlorite 0.1 - 1%)	According to Manufacturer Spec	11-Aug-16
Spartan Chemical Company Inc.	BNC-15™	According to Manufacturer Spec	11-Aug-16
Ameriderm Laboratories Limited	Care Source Disinfectant	According to Manufacturer Spec	10-Nov-11
Metrex® Research	CaviCide®	According to Manufacturer Spec	06-Oct-16
Metrex® Research	CaviCide1™	According to Manufacturer Spec	05-Jan-17
Metrex® Research	CaviWipes1™	Ready to Use	05-Jan-17

PLEASE REFER TO IMPORTANT MESSAGES ON FINAL PAGE BEFORE APPLICATION OF ANY CLEANER/DISINFECTANT.

Sorted by Product Name:

MANUFACTURER	PRODUCT	CONCENTRATION	DATE TESTED
Metrex® Research	CaviWipes™	Ready to Use	06-Oct-16
Helix Solutions	Chlor-Clean	1000PPM (1 Tablet per 1 Litre Water)	10-Oct-11
Agar Cleaning Systems Pty. Ltd.	Chloradet	Minimum 160 to 1 dilution with water	06-Jun-16
Ecolab®	Chlorwhite	Minimum 100 to 1 dilution with water	22-Oct-13
Apollo Corporation	CID A-L™ II Quaternary Disinfectant	According to Manufacturer Spec	26-Sep-12
Wexford Labs, Inc.	CleanCide® Wipes	Ready to Use	15-May-19
Clorox Disinfecting Wipes	Clorox Disinfecting Wipes	Ready to Use	14-Feb-11
Alco Mandala Products	DC-7	According to Manufacturer Spec	16-Feb-12
Delaval	Delaval Acid Wash	Minimum 20 to 1 dilution with water	06-Jun-14
Clinell®	Detergent Wipes	According to Manufacturer Spec	27-May-09
ArjoHuntleigh	Disinfectant Cleanser IV	According to Manufacturer Spec	23-Nov-09
Clorox®	Dispatch® Disinfectant	According to Manufacturer Spec	16-Feb-12
Tristel Solutions LTD	Distel Surface Disinfectant (Formerly Trigene®)	According to Manufacturer Spec	24-Oct-06
Sealed Air Diversey Care	Divercleanse	Minimum 33 to 1 dilution with water	15-Dec-09
Buckeye® International, Inc.	Eco E23 S23	Minimum 256 to 1 dilution with water	05-Jan-17
Velocity Chemicals	Eco-Pine	Minimum 20 to 1 dilution with water	11-Apr-06
Clorox HEALTHCARE®	Fuzion™	According to Manufacturer Spec	05-Jan-17

CARR NORTH AMERICA LLC

Toll Free: 800.875.5677

540 Kenneth Welch Drive, Lakeville, MA 02347, U.S.A.

PO Box 988, Lakeville, MA 02347, U.S.A.

info@carrnorthamerica.com

PLEASE REFER TO IMPORTANT MESSAGES ON FINAL PAGE BEFORE APPLICATION OF ANY CLEANER/DISINFECTANT.

Sorted by Product Name:

MANUFACTURER	PRODUCT	CONCENTRATION	DATE TESTED
Spartan Chemical Company Inc.	GS Neutral Disinfectant #3502	According to Manufacturer Spec	02-Dec-14
Clinell®	Hand & Surface Sanitizer	According to Manufacturer Spec	27-May-09
3M	HB Quat Disinfectant	Ready to Use	06-Sep-19
Spartan Chemical Company Inc.	hdq®C2	According to Manufacturer Spec	30-Jun-17
Brighton Professional	Hepastat 256	According to Manufacturer Spec	27-Aug-14
Wipesplus	Hospital/Surgical Wipes	Ready to Use	15-May-19
(Various)	Hydrogen Peroxide 5%	Ready to Use	25-May-12
Clorox HEALTHCARE®	Hydrogen Peroxide Cleaner Disinfectant Wipes	Ready to Use	12-Apr-19
Apollo Corporation	Hygena™	According to Manufacturer Spec	25-Sep-12
Ecolab®	Incidin® Extra	According to Manufacturer Spec	06-Jun-14
Ecolab®	Incidin® Foam	According to Manufacturer Spec	06-Jun-14
Ecolab®	Incidin® Plus	According to Manufacturer Spec	06-Jun-14
(Various)	Isopropyl Alcohol 70%	Ready to Use	25-Feb-10
Kimberly-Clark®	Isowipes®	Ready to Use	19-Jul-12
Schulke & Mayr GMBH	Kodan® Tinkture Forte	According to Manufacturer Spec	06-Jun-14
Legge Systems	Legcide HB Plus	According to Manufacturer Spec	*****
Whiteley Medical	Lemex	According to Manufacturer Spec	15-Mar-11

PLEASE REFER TO IMPORTANT MESSAGES ON FINAL PAGE BEFORE APPLICATION OF ANY CLEANER/DISINFECTANT.

Sorted by Product Name:

MANUFACTURER	PRODUCT	CONCENTRATION	DATE TESTED
Livingstone International (PTY) LTD	LIV-Wipes Jumbo	Ready to Use	15-Mar-11
Sulco LTD	Mediwipes	Ready to Use	27-May-09
Medline Industries Inc.	MICRO-KILL™ + Disinfecting, Deodorizing, Cleaning Wipes	According to Manufacturer Spec	09-Mar-20
Medline Industries Inc.	MICRO-KILL™ AF2 Wipes	Ready to Use	15-May-19
Medline Industries Inc.	MICRO-KILL™ Bleach Germicidal Wipes	Ready to Use	15-May-19
Medline Industries Inc.	MICRO-KILL™ Bleach-Free, Alcohol-Free Wipes	Ready to Use	15-May-19
Medline Industries Inc.	MICRO-KILL™ One Wipes	Ready to Use	15-May-19
Medline Industries Inc.	MICRO-KILL™ Q3 Concentrated Disinfectant, Cleaner & Deodorizer	According to Manufacturer Spec	01-Feb-18
Pro Restore® Products	Microban X-590 Institutional Spray	According to Manufacturer Spec	11-Aug-16
Anti-Slip Anywhere	Multi Kleen	According to Manufacturer Spec	15-Mar-11
Medizar®	Multi-Surface Sanitiser Wipes	Ready to Use	17-May-19
Ecolab® Cleantec	Nature Clean	Minimum 38 to 1 dilution with water	17-Dec-10
Certo®	Non Acid Bathroom Cleaner	According to Manufacturer Spec	07-May-12
PreVasive®	Noroxycdiff™	According to Manufacturer Spec	22-May-19
Ecolab®	Oasis Pro 12 Neutral Cleaner	Minimum 20 to 1 dilution with water	06-Mar-12
Schülke & Mayr GmbH	Ocetenisept®	According to Manufacturer Spec	06-Jun-14
Whiteley Medical	Odacon Spray	According to Manufacturer Spec	14-Jun-11

PLEASE REFER TO IMPORTANT MESSAGES ON FINAL PAGE BEFORE APPLICATION OF ANY CLEANER/DISINFECTANT.

Sorted by Product Name:

MANUFACTURER	PRODUCT	CONCENTRATION	DATE TESTED
OdorXit Naturals	Odor Eliminator Spray	Ready to Use	22-Feb-18
Sealed Air Diversey Care	Oxiver Five 16	Minimum 16 to 1 dilution with water	08-Nov-19
Sealed Air Diversey Care	Oxiver One Wipes	Ready to Use	14-Feb-20
Sealed Air Diversey Care	Oxivir® TB Liquid and Wipes	Ready to Use	22-Apr-13
Ecolab®	Oxycide	According to Manufacturer Spec	06-Jul-14
Pal® International	PAL® Detergent Wipes	Ready to Use	27-May-09
Pal® International	PAL® Disinfectant Wipes	Ready to Use	27-May-09
PCS Process Cleaning Solutions	PCS 7000	According to Manufacturer Spec	06-Oct-17
PCS Process Cleaning Solutions	PCS Spray & Wipe Hand and Surface Cleaner	Ready to Use	06-Oct-17
Pure Green, LLC	Puregreen 24 (Pure Green)	According to Manufacturer Spec	30-Jan-09
Reynard Health Supplies	Reynard Neutral Detergent Wipes	Ready to Use	08-Feb-07
Reynard Health Supplies	Reynard Surface Sanitizing Wipes	Ready to Use	27-May-11
Sealed Air Diversey Care	Room Care R2L	According to Manufacturer Spec	24-Oct-12
STERI-7®	S-7 Xtra Wipes	Ready to Use	28-Nov-16
Buckeye® International, Inc.	Sanicare TBX™	According to Manufacturer Spec	21-Sep-10
Buckeye® International, Inc.	Sanicare™ Disinfectant Wipes	Ready to Use	21-Sep-10
Alco Mandala Products	Sanni-Rinse	According to Manufacturer Spec	16-Feb-12

PLEASE REFER TO IMPORTANT MESSAGES ON FINAL PAGE BEFORE APPLICATION OF ANY CLEANER/DISINFECTANT.

Sorted by Product Name:

MANUFACTURER	PRODUCT	CONCENTRATION	DATE TESTED
Hu-Friedy Mfg. Co., LLC.	Sporeclear RTU Spray or Wipes	Minimum 13 to 1 dilution with water	31-Jan-16
Clinell®	Sporicidal Wipes	Ready to Use	27-May-09
Ecolab®	Spring Clean	Minimum 20 to 1 dilution with water	11-Apr-06
Noble Pine Products Co.	Steri-Fab®	According to Manufacturer Spec	11-Aug-16
Hartmann	Sterillium®	According to Manufacturer Spec	06-Jun-14
sBioMed, LLC	Steriplex SD Disinfectant Sporicide Wipes (Without Bleach)	According to Manufacturer Spec	20-May-15
Sterri-Matt®	Sterri-Matt® (Pure) Foam	Ready to Use	15-Dec-09
Sealed Air Diversey Care	Stride® Citrus Smartdose	Minimum 750 to 1 dilution with water	01-Jun-15
Sealed Air Diversey Care	Stride® HC (JF)	According to Manufacturer Spec	01-Jun-15
Sealed Air Diversey Care	Suma San Conc D4A	According to Manufacturer Spec	28-Oct-14
Sealed Air Diversey Care	Suma San Conc D4A	According to Manufacturer Spec	28-Oct-14
Spartan Chemical Company Inc.	Super HDQ Neutral®	Minimum 256 to 1 dilution with water	11-Aug-16
PDI	Super Sani Cloth® Germicidal Disposable Wipes	Ready to Use	02-Sep-10
Anios Laboratories	Surfa 'Safe	According to Manufacturer Spec	06-Jun-14
Bioscience Fresche®	Surface Plus	According to Manufacturer Spec	09-Feb-15
Reynard Health Supplies	Surface Sanitising Wipes	Ready to Use	18-Feb-09
3M	TB Quat Disinfectant	Ready to Use	09-Mar-18

PLEASE REFER TO IMPORTANT MESSAGES ON FINAL PAGE BEFORE APPLICATION OF ANY CLEANER/DISINFECTANT.

Sorted by Product Name:

MANUFACTURER	PRODUCT	CONCENTRATION	DATE TESTED
Buckeye® International, Inc.	Terminator™ One Step Disinfectant	According to Manufacturer Spec	21-Sep-10
Apollo Corporation	Therasol™	According to Manufacturer Spec	25-Sep-12
Medline Industries Inc.	Touch Screen Cleaning Wipes	Ready to Use	15-May-19
Hunters Products	TRI-DET	Minimum 80 to 1 dilution with water	02-Dec-14
(Various)	Trix (Soap)	Ready to Use	27-Apr-10
Verna Care	Tuffie 5 Cleaning Wipes	Ready to Use	19-Jul-12
Apollo Corporation	Turbo Clean™	According to Manufacturer Spec	25-Sep-12
Texwipe®	TX®3270	Ready to Use	12-Apr-19
Clinell®	Universal Disinfectant Wipes	According to Manufacturer Spec	27-May-09
Clinell®	Universal Wipes	Ready to Use	12-May-19
True Blue Chemicals	Use All	According to Manufacturer Spec	28-Sep-12
Whiteley Medical	Viraclean®	According to Manufacturer Spec	19-Jul-12
Sealed Air Diversey Care	Virex® II 256	Minimum 256 to 1 dilution with water	07-May-12
Sealed Air Diversey Care	Virex® Plus	According to Manufacturer Spec	19-Dec-18
Dupont™	Virkon®	Minimum 20 to 1 dilution with water	15-Dec-09
Wexford Labs	Wexcide 128	Minimum 128 to 1 dilution with water	11-Aug-16

CAUTION!

To avoid premature degradation of the surface coating never use an abrasive cloth, sponge or other similar device on the surface. Carefully follow manufacturer instructions for recommended exposure time of cleaner on surface. Wipe off cleaning product residue with a clean, soft cloth after cleaning and allow the surface to dry completely before use.

IMPORTANT INFORMATION:

Cleaners/disinfectants can cause discoloration/bleaching on some colors. To avoid discoloration, always check on a small area first.

Information contained here is purely related to the compatibility of cleaners with CARRFLEX® materials and in no way relates to the efficacy of any cleaner for infection control or the safety of any cleaner for skin contact.

Testing info is specific to the products or formulations listed and in no way constitutes approval for other products from any brand listed without being tested.

All approvals are based on manufacturer recommended dilution with water unless otherwise specified and approval is void for any improper dilution.

Never mix different cleaners. Results can be deadly.

If you wish to change from one approved cleaner to another, thoroughly rinse the surface with warm water, allow to dry completely and repeat this process one more time prior to use of another cleaner.

If a particular cleaner/disinfectant is not listed, we advise to test the specific product in advance or to contact us to conduct testing.

Listed chlorine based cleaners should be used only in instances of irregular spot cleaning and bodily fluid spills.

CARR NORTH AMERICA LLC

Toll Free: 800.875.5677

540 Kenneth Welch Drive, Lakeville, MA 02347, U.S.A.

PO Box 988, Lakeville, MA 02347, U.S.A.

info@carrnorthamerica.com